



Specific Accreditation Criteria Life Sciences ISO/IEC 17025 Annex

Facilities testing for genetically modified organisms (GMO)

January 2018



© Copyright National Association of Testing Authorities, Australia 2013


This publication is protected by copyright under the Commonwealth of Australia Copyright Act 1968.

NATA's accredited facilities or facilities seeking accreditation may use or copy this publication or print or email this publication internally for accreditation purposes.

Individuals may store a copy of this publication for private non-commercial use or copy a reasonable portion of this publication in accordance with the fair dealing provisions in Part III Division 3 of the Copyright Act 1968.

You must include this copyright notice in its complete form if you make a copy of this publication.

Apart from these permitted uses, you must not modify, copy, reproduce, republish, frame, upload to a third party, store in a retrieval system, post, transmit or distribute this content in any way or any form or by any means without express written authority from NATA.



Facilities testing for genetically modified organisms (GMO)

This document provides additional interpretative criteria and recommendations for the application of ISO/IEC 17025 for both applicant and accredited facilities conducting testing for GM material.

Applicant and accredited facilities must also comply with ISO/IEC 17025, the NATA ISO/IEC 17025 Standard Application Document (SAD) and the Life Sciences ISO/IEC 17025 Appendix.

The clause numbers in this document follow those of ISO/IEC 17025 but since not all clauses require interpretation the numbering may not be consecutive.

Testing for genetically modified sequences in foodstuffs or whole grain and related plant materials relies on the ability to test for the specific DNA sequences associated with modifications and/or the promoting and terminating sequences associated with the inserted sequences.

Accreditation covers qualitative and quantitative analysis of GM material through the use of DNA extraction and Polymerase Chain Reaction (PCR) methods.

In recognition of the continual development of new GM materials NATA operates a flexible scope policy. When a facility has developed the capacity to detect a new GM event or construct this will be added to the scope of accreditation without the need for supporting documentation to be submitted to NATA where the analysis is undertaken using the same technique that is covered by the scope of accreditation. Where a new technique is implemented that has not been assessed an on site variation visit will be required. The competence of a facility to develop and implement tests for new GM events and constructs will be ascertained at the initial assessment. At reassessment a sample of GM events or constructs added since the last technical assessment will be reviewed.

5.3 Accommodation and environmental conditions

In order to reduce the risk of false positive results from cross-contamination or carry-over contamination of samples and reagents by other samples in the laboratory or by amplified material, four physically separate and contained areas with known air conditioning/ventilation airflows are required within a facility undertaking nucleic acid amplification for the detection of GM materials. The four work areas are:

- Sample preparation
- DNA extraction
- Reagent preparation
- Product analysis

5.4 Test and calibration methods and method validation

The following reference documents covering nucleic acid testing should be consulted:

- Former Subcommittee of Animal Health Laboratory Standards (SCAHLs) *Veterinary Laboratory Guidelines for Nucleic Acid Detection Techniques*
- National Pathology Accreditation Advisory Council *Laboratory Requirements for medical testing of human nucleic acids 2013*

5.4.5 Method validation

Where possible consideration should be given to grouping like matrices to reduce the volume of work required to undertake validation.

See the FSANZ website at www.foodstandards.gov.au information on the currently approved/pending GM events.

5.5 Equipment

Specialised equipment (i.e. thermocycler, nucleic acid analyser, spectrophotometer) is required for detection of labeled DNA fragments. Many instruments have internal diagnostic checks built into them and, as a minimum, the recommendations from manufacturers should be followed.

5.6 Measurement Traceability

Reference materials used as controls in the validation of new assays and as routine controls must comply with the requirements of NATA's *Metrological Traceability*.

5.8 Handling of test and calibration items

The high sensitivity of methods dictates a higher than usual awareness to the possibility of cross-contamination during transport, storage, preparation and analysis. (See 5.3 Accommodation.)

5.9 Assuring the quality of test and calibration results

Controls that are incorporated with the GMO test system must address inhibition, sensitivity and contamination.

5.10 Reporting of results

Reporting the results of GM material detection must be sufficiently descriptive to allow the reader a clear understanding of what tests have been conducted and to what level of detection.

In addition to the general requirements of ISO/IEC 17025, a test report for GM material testing must also include information on the limitations of the particular detection method used for the particular sample or sample type:

The content of test reports is influenced by the nature of the matrix tested. For unprocessed materials such as soya beans, canola seeds or lupin seeds it is important to include the amount of sample or the number of seeds ground prior to DNA extraction.

Where highly processed materials are tested it is possible that an insufficient quantity of DNA can be extracted to allow the presence of GM material to be detected. In such cases a null result can be reported. Facilities should discuss the possibility of a null result with customers prior to testing if the matrix submitted is highly processed.

No affirmation shall be made stating that there is no GM material present in the sample analysed as determined from test samples.

References

Veterinary Laboratory Guidelines for Nucleic Acid Detection Techniques 2008
Former Subcommittee of Animal Health Laboratory Standards

Requirements for medical testing of human nucleic acids 2013
National Pathology Accreditation Advisory Council

NATA Publications

Life Sciences ISO IEC 17025 Appendix

Amendment Table

The table below provides a summary of changes made to the document with this issue.

Section or Clause	Amendment
New document	This document represents a direct adoption of the former Biological Testing Annex D. The document has been reviewed and updated to reflect the new accreditation criteria documentation structure.