



# **Specific Accreditation Criteria**

## **ISO/IEC 17025 Application Document Calibration - Annex**

### **Dimensional metrology**

**Issued: May 2019**

**Effective: August 2019**

**© Copyright National Association of Testing Authorities, Australia 2014**

This publication is protected by copyright under the Commonwealth of Australia Copyright Act 1968.

NATA's accredited facilities or facilities seeking accreditation may use or copy this publication or print or email this publication internally for accreditation purposes.

Individuals may store a copy of this publication for private non-commercial use or copy a reasonable portion of this publication in accordance with the fair dealing provisions in Part III Division 3 of the Copyright Act 1968.

You must include this copyright notice in its complete form if you make a copy of this publication.

Apart from these permitted uses, you must not modify, copy, reproduce, republish, frame, upload to a third party, store in a retrieval system, post, transmit or distribute this content in any way or any form or by any means without express written authority from NATA.

# Table of Contents

Purpose..... 4

Criteria and recommendations applicable to all types  
of dimensional measurements ..... 4

Ultrasonic thickness gauges..... 4

References..... 5

Amendment table ..... 5

## Dimensional metrology

### Purpose

In addition to the *ISO/IEC 17025 Standard Application Document (SAD)* and the accompanying Calibration - Appendix, this document provides interpretative criteria and recommendations for dimensional metrology for both applicant and accredited facilities.

Facilities must comply with all relevant documents in the NATA Accreditation Criteria (NAC) package for Calibration (refer to *NATA Procedures for Accreditation*).

For ease of use and to avoid fragmentation of the information, the relevant clause numbers of ISO/IEC 17025 have not been included.

### Criteria and recommendations applicable to all types of dimensional measurements

Facilities must be familiar with the filtering characteristics of the reference instruments they use. The potential loss or distortion of captured information must be considered when selecting filter settings as well as their effect on any time-related phenomena.

Records of these settings must be retained and/or be specified in the calibration or testing procedures.

In roundness measurement, significant differences in results can occur on test items with certain irregularities depending on the filter type and cut-off value selected. Facilities should normally default to a low level of filtering for high quality surfaces (e.g. 1:500 UPR).

Ideally, all measurements will be carried out under static conditions, however in some force measurements where test machines have limited control or creep effects are occurring, different filtering (indicator averaging and update rates) used on the test and reference instruments can introduce errors into the measurements.

### Ultrasonic thickness gauges

As accuracy of an ultrasonic thickness gauge is highly related to the material to be measured and the speed of sound in this material, the calibration report must include the information regarding the reference material and the speed of sound applied.

For example:

*The thickness gauge was calibrated using stainless steel gauge blocks and the speed of sound set at 5790 m/s. The temperature during calibration was 20°C ±1°C.*

## References

This section lists publications referenced in this document. The year of publication is not included as it is expected that only current versions of the references shall be used.

### Standards

ISO/IEC 17025      General requirements for the competence of testing and calibration laboratories

### NATA Publications

NATA Accreditation Criteria (NAC) package for Calibration

General Accreditation Criteria      ISO/IEC 17025 Standard Application Document

Specific Accreditation Criteria      ISO/IEC 17025 Standard Application Document, Calibration - Appendix

## Amendment table

The table below provides a summary of changes made to the document with this issue.

Section or Clause	Amendment
Ultrasonic thickness gauges	New section covering reporting criteria for these devices.
Whole document	Addition of Security Classification Label